

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

RM
24-6486

Date: 02/13/2024

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:49
INTERNAL STANDARD	VERIFIED	14:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:50
BLANK TEST	0.00	14:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:51
BLANK TEST	0.00	14:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:52
BLANK TEST	0.00	14:53

