

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

Date: 01/09/2024

*KMS*  
*24-1289*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:49
INTERNAL STANDARD	VERIFIED	09:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:49
BLANK TEST	0.00	09:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:51
BLANK TEST	0.00	09:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:52
BLANK TEST	0.00	09:53

