

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 02/13/2024

*LMJ*  
*29-6486*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:42
INTERNAL STANDARD	VERIFIED	13:42
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	13:43
BLANK TEST	0.00	13:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:44
BLANK TEST	0.00	13:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:45
BLANK TEST	0.00	13:46

