

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 04/02/2024

*LMS*  
*24-14395*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:35
INTERNAL STANDARD	VERIFIED	12:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:36
BLANK TEST	0.00	12:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:37
BLANK TEST	0.00	12:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:38
BLANK TEST	0.00	12:39

