

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 04/16/2024

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24040

LMS
29-16667

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:13
INTERNAL STANDARD	VERIFIED	14:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:14
BLANK TEST	0.00	14:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:15
BLANK TEST	0.00	14:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:16
BLANK TEST	0.00	14:17

